



eTimestrip Report

Report Generated At:26-Oct-2022
15:03:20



Device Information

Device Code:	TL520	Type:	Temperature
Serial Number:	TLA220300300	Version:	V2.3
Time Zone:	UTC +00:00	Working status:	Record

Trip Information

Description:

Configuration Information

Start Type:	Break Tab	Start Delay:	0H 30M
Sampling Interval:	0H 1M	Logging Interval:	0H 5M
Data categories:	Avg	Data Storage Type:	Temp
Led Flashing:	30 Seconds		

Alarm Limit	Alarm Delay	Alarm Type	Total Time	Started	Violations	Status
T1 :Higher 8.0°C	0H 30M	Single	0H 26M	--	1	OK
T0 :Below 2.0°C	0H 30M	Single	0H 0M	--	0	OK

Alarm Zone	Duration Of The First Overtemp	Statistics Of The First Overtemp
T1 :Higher 8.0°C	0H 26M	20.6°C(Avg)
T0 :Below 2.0°C	0H 0M	0.0°C(Avg)

Summary




Max:	25.5°C	Min:	4.7°C
------	--------	------	-------

Number of Samples:	186	Number of Records:	38
--------------------	-----	--------------------	----

Start Time:	26-Oct-2022 02:57:44	Last Sampling Time:	26-Oct-2022 06:02:44(UTC +00:00)
-------------	-------------------------	---------------------	--

Running Time:	3H 5M
---------------	-------



Time	Temp(°C)	Time	Temp(°C)	Time	Temp(°C)	Time	Temp(°C)
2022-10-26	5.3						
03:02:44	5.4						
03:07:44	5.0						
03:12:44	5.7						
03:17:44	5.3						
03:22:44	5.4						
03:27:44	4.9						
03:32:44	5.1						
03:37:44	4.8						
03:42:44	5.0						
03:47:44	4.7						
03:52:44	5.0						
03:57:44	4.7						
04:02:44	5.0						
04:07:44	5.2						
04:12:44	5.5						
04:17:44	5.6						
04:22:44	6.3						
04:27:44	6.4						
04:32:44	6.4						
04:37:44	6.5						
04:42:44	6.5						
04:47:44	6.6						
04:52:44	6.7						
04:57:44	6.7						
05:02:44	6.7						
05:07:44	6.8						
05:12:44	6.9						
05:17:44	6.9						
05:22:44	7.1						
05:27:44	6.8						
05:32:44	5.8						
05:37:44	8.3						
05:42:44	24.5						
05:47:44	23.4						
05:52:44	22.8						
05:57:44	22.9						
06:02:44	9.7						