



the smart way to measure time & temperature

# Certificate of Calibration

PRODUCT TL520

STATEMENT OF CALIBRATION The devices with serial numbers listed below have been thoroughly tested, validated and meet performance accuracy specifications over the stated ranges. The certificate is valid for three years from the date of calibration.

REFERENCE EQUIPMENT The reference instruments have been calibrated by the China National Accreditation Council for Conformity Assessment and can be traced back to the National Institute of Standards and Technology (NIST), USA under multiple accreditation agreements.

TEMP ACCURACY Temp Accuracy  
-20~+40°C(±0.5°C)

Others  
±1.0°C

VALIDATION POINTS	INDICATED POINTS	VARIANCE	TOLERANCE	RESULT
+60°C	+59.1°C	-0.9°C	±1.0°C	PASS
+40°C	+39.7°C	-0.3°C	±0.5°C	PASS
+20°C	+20.2°C	+0.2°C	±0.5°C	PASS
+8°C	+7.7°C	-0.3°C	±0.5°C	PASS
+2°C	+2.3°C	+0.3°C	±0.5°C	PASS
-20°C	-20.4°C	-0.4°C	±0.5°C	PASS

Timestrip UK Ltd Sheraton House, Castle Park, Cambridge, CB3 OAX United Kingdom

[www.timestrip.com](http://www.timestrip.com)

SERIAL NUMBER TLA220300201 - TLA220300300

DATE OF CALIBRATION Mar 2022

INSPECTED BY:

*Jack Jia*